

CALIBRATION CERTIFICATE

1709-12764

Customer information

Client : Heckmann Präzisionstechnik GmbH (Digi-Pas)
 Contact : Fr. Sarah Heckmann
 Address : Wilhelm-Leuschner-Str. 36
 68519 Viernheim
 Germany
 Reference client :
 Reference Trescal : 201718045/2

Instrument information

Make / type : DIGI-PAS / DWL-1500 XY
 Description : Inclinometer
 Range : 0 .. 2 °
 Serial number : 11A21763
 Identification number :
 Accuracy :

Date of calibration : 16 October 2017 to 11 January 2018

Method of calibration

P1-02-G.017 Calibration of spirit levels / inclinometers

The calibration of the levels/clinometers consists of a visual inspection and a series of measurements. Firstly, we examine the general condition of the surfaces and the functionality of the readout. Nextly we measure the zero, the deviations and, if necessary, the perpendicularity of the instrument.

Environmental conditions (limits during measurements)

Ambient temperature : 20 °C ± 1 °C
 Relative humidity : 45%rh ± 20%rh

Used reference

The equipment used is traceable to National and/or International standards.
 R3797/3 Rotary encoder Cert.170504038

Note

The instrument is measured, adjusted and then re-measured.

Issue date: 12 January 2018

Technician
Koen Groffen



Head of the laboratory
Luc Van Pelt



BELAC is member of the European Co-operation for Accreditation (EA) and is one of the signatories of the EA Multilateral Agreement and to the ILAC (International Laboratory Accreditation Co-operation) Mutual Recognition Arrangements (MRA) for the mutual recognition of calibration certificates.
 This document is issued in accordance with the conditions for accreditation of the BELAC which is based on ISO/IEC 17025.
 This document may not be reproduced other than in full, except with the prior written approval of the head of the issuing laboratory.
 Unless otherwise stated, the calibration was performed at the address mentioned in the footnote.

Trescal nv | Vosstraat 200 | 2600 Berchem (Antwerpen) | Belgium | T +32 3 542 62 90 | E info.benelux@trescal.com



* C 1 7 0 9 1 2 7 6 4 *

CALIBRATION CERTIFICATE

1709-12764

Initial results:

-	Description	Reference value	Instrument value	Difference	Tolerance ±	Uncertainty ±	Units	-
1	◀	0,000	0,000	0,000	0,001	0,001	°	
2	◀	0,200	0,202	0,002	0,001	0,001	°	*
3	◀	0,500	0,503	0,003	0,001	0,001	°	*
4	◀	1,000	1,004	0,004	0,003	0,001	°	*
5	◀	1,500	1,505	0,005	0,003	0,001	°	*
6	▶	0,000	0,000	0,000	0,001	0,001	°	
7	▶	0,200	0,200	0,000	0,001	0,001	°	
8	▶	0,500	0,500	0,000	0,001	0,001	°	
9	▶	1,000	1,003	0,003	0,003	0,001	°	
10	▶	1,500	1,504	0,004	0,003	0,001	°	*

-	Description	Reference value	Instrument value	Difference	Tolerance ±	Uncertainty ±	Units	-
1	▼	0,000	0,000	0,000	0,001	0,001	°	
2	▼	0,200	0,201	0,001	0,001	0,001	°	
3	▼	0,500	0,502	0,002	0,001	0,001	°	*
4	▼	1,000	1,004	0,004	0,001	0,001	°	*
5	▼	1,500	1,505	0,005	0,001	0,001	°	*
6	▲	0,000	0,000	0,000	0,001	0,001	°	
7	▲	0,200	0,200	0,000	0,001	0,001	°	
8	▲	0,500	0,502	0,002	0,003	0,001	°	
9	▲	1,000	1,003	0,003	0,003	0,001	°	
10	▲	1,500	1,506	0,006	0,003	0,001	°	*

After adjustment results:

-	Description	Reference value	Instrument value	Difference	Tolerance ±	Uncertainty ±	Units	-
1	◀	0,000	0,000	0,000	0,001	0,001	°	
2	◀	0,200	0,201	0,001	0,001	0,001	°	
3	◀	0,500	0,500	0,000	0,001	0,001	°	
4	◀	1,000	1,000	0,000	0,001	0,001	°	
5	◀	1,500	1,499	-0,001	0,001	0,001	°	
6	▶	0,000	0,000	0,000	0,001	0,001	°	
7	▶	0,200	0,200	0,000	0,001	0,001	°	
8	▶	0,500	0,500	0,000	0,003	0,001	°	
9	▶	1,000	1,000	0,000	0,003	0,001	°	
10	▶	1,500	1,500	0,000	0,003	0,001	°	

CALIBRATION CERTIFICATE

1709-12764

-	Description	Reference value	Instrument value	Difference	Tolerance ±	Uncertainty ±	Units	-
1	▼	0,000	0,000	0,000	0,001	0,001	°	
2	▼	0,200	0,200	0,000	0,001	0,001	°	
3	▼	0,500	0,500	0,000	0,001	0,001	°	
4	▼	1,000	0,999	-0,001	0,001	0,001	°	
5	▼	1,500	1,500	0,000	0,001	0,001	°	
6	▲	0,000	0,000	0,000	0,001	0,001	°	
7	▲	0,200	0,200	0,000	0,001	0,001	°	
8	▲	0,500	0,500	0,000	0,003	0,001	°	
9	▲	1,000	0,999	-0,001	0,003	0,001	°	
10	▲	1,500	1,500	0,000	0,003	0,001	°	

The stated uncertainty is that of the entire set-up including the object under test.

The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95% .

The uncertainty is calculated following EA-4/02 in accordance with the requirements of the ISO/IEC 17025.